

Determination of *N,N*-Dimethyl-*o*-Toluidine and *N,N*-Diethyl-*p*-Toluidine in Ethylene Gas Samples

INTRODUCTION

Amines are found throughout nature and are synthesized for many commercial applications. They are used as pharmaceuticals, anti-corrosive agents, and chemical intermediates, and can indicate food spoilage. Ion chromatography (IC) with either suppressed conductivity or integrated amperometry detection is an excellent technique for determining amines in a variety of samples, even at low concentrations.¹⁻³ Dionex has designed high-performance cation-exchange columns, the IonPac® CS17 and IonPac CS18, to separate amines from common inorganic cations and ammonium and deliver good peak shapes. Although these columns can be used for a number of amines with a weak (< 50 mM) methanesulfonic acid (MSA) eluent, some amines do require the addition of an organic solvent, usually acetonitrile, to the eluent to achieve good peak shape and shorten analysis time. There are also some amines that can be detected successfully by UV absorbance. Here is shown the separation and detection of two such amines: *N,N*-dimethyl-*o*-toluidine and *N,N*-diethyl-*p*-toluidine. These amines are

separated on a CS17 column with an MSA gradient and 9% acetonitrile, and are detected by absorbance at 210 nm. This method was used to monitor *N,N*-dimethyl-*o*-toluidine and *N,N*-diethyl-*p*-toluidine in ethylene gas. The gas was bubbled into a solution of MSA to trap the amine impurities.

EXPERIMENTAL

Equipment

Dionex ICS-3000 Reagent-Free™ Ion Chromatography system with:

DP Dual Pump module^a

DC Detector/Chromatography module

AS Autosampler^b

PDA-100 Photodiode Array Detector^c

Chromeleon® 6.8 Chromatography Management Software

^aFor this application, a Single Pump (SP) module can also be used.

^bAn AS40 Autosampler can also be used for this application.

^cA VWD Variable Wavelength Detector can also be used for this application if spectral identification is not needed.

Now sold under the
Thermo Scientific brand

Thermo
SCIENTIFIC

Reagents and Standards

Deionized (DI) water, Type I reagent grade, 18 M $\cdot$ cm resistivity or better

N,N-dimethyl-*o*-toluidine (C₉H₁₃N)

N,N-diethyl-*p*-toluidine (C₁₁H₁₇N)

Methanesulfonic acid (MSA) (CH₄O₃S)

Acetonitrile (CH₃CN)

All compounds were ACS reagent grade or better from reliable sources. The *N,N*-dimethyl-*o*-toluidine (C₉H₁₃N) and *N,N*-diethyl-*p*-toluidine (C₁₁H₁₇N) were gifts from a Dionex customer, but *N,N*-dimethyl-*o*-toluidine can be obtained from Sigma-Aldrich (St. Louis, MO, USA) and both compounds from R.S.A. Corporation (Danbury, CN, USA).

Conditions

Column: IonPac CS17 2 $\times$ 250 mm (P/N 060561)*

Eluents: (E1) DI water
(E2) 100 mM Methanesulfonic acid
(E3) 90% Acetonitrile

Gradient Steps:	Time (min)	%E1	%E2	%E3
	-7.0	86	4	10
	0.0	86	4	10 (inject)
	3.5	86	4	10
	12.0	10	80	10
	15.9	10	80	10
	16.0	86	4	10

Flow Rate: 0.4 mL/min

Inj. Volume: 100 μ L

Column Oven: 40 $^{\circ}$ C

System

Backpressure: 2400-2800 psi

Detection: UV at 210 nm (PDA mode)

*A guard column was not used for this application in order to keep the backpressure under 3000 psi. The system and column can accommodate higher backpressures but all fittings must be tight to avoid leaks.

Preparation of Reagents

Methanesulfonic acid (MSA) 100 mM

Dissolve 9.610 g of MSA in a 1 L volumetric flask containing approximately 300 mL of deionized water. Bring to volume with deionized water, mix well, degas, and transfer to an eluent bottle.

90% Acetonitrile

Mix 900 mL of acetonitrile and 100 mL of deionized water. Mix well and degas.

Preparation of Standard Solutions

Stock Standards

Prepare 1000 mg/L standards for each of the analytes in eluent (4 mM MSA, 9 % acetonitrile). Standards should be prepared from the highest purity compounds available. Table 1 provides the amounts needed to prepare 100 mL of each standard. Concentrated standards are stable for at least one month when stored at 4 $^{\circ}$ C.

Table 1. Masses of Compounds Used to Prepare 100 mL of 1000 mg/L Amine Standards

Analyte	Compound	Amount (g)
<i>N,N</i> -dimethyl- <i>o</i> -toluidine	<i>N,N</i> -dimethyl- <i>o</i> -toluidine	0.100
<i>N,N</i> -diethyl- <i>p</i> -toluidine	<i>N,N</i> -diethyl- <i>p</i> -toluidine	0.100

Secondary Standards

The stock standards are used to prepare the calibration standards listed in Table 2. For example, the level 3 standard is prepared by adding 1 mL each of the 1000 mg/L stock standards of *N,N*-dimethyl-*o*-toluidine and *N,N*-diethyl-*p*-toluidine to a 100 mL volumetric flask and then bringing to volume with eluent (4 mM MSA, 9 % acetonitrile).

Table 2. Standard Concentrations for Method Calibration

Analyte	Concentration (mg/L)		
	Level 1	Level 2	Level 3
<i>N,N</i> -dimethyl- <i>o</i> -toluidine	1	5	10
<i>N,N</i> -diethyl- <i>p</i> -toluidine	1	5	10




Figure 1. Chromatography of the MSA blank.

Preparation of Samples

A Milligan gas washing tower (scrubber) was charged with 100 mL of 25 mM MSA solution. Prior to bubbling the gas through the MSA solution in the Milligan bottle, a sample was removed to serve as a blank. To trap the amines, the ethylene gas was bubbled (scrubbed) through the bottle at a flow rate of 200 mL/min (steady stream of bubbles) until the sample cylinder was empty. These samples were kindly provided by a customer.

RESULTS AND DISCUSSION

Figures 1 and 2 show the results of the method development for the separation of *N,N*-dimethyl-*o*-toluidine and *N,N*-diethyl-*p*-toluidine on the IonPac CS17 column. The separation uses a short isocratic 4 mM MSA eluent followed by a gradient from 4 to 80 mM MSA all in the presence of 9% acetonitrile. The acetonitrile is added to ensure the best peak shapes for these hydrophobic amines. Figure 2 shows the baseline separation of the amines at three concentrations. While wavelength scans of these amines (Figures 3 and 4) reveal that they can be detected at approximately 260 nm, 210 nm was chosen for the best sensitivity. The choice of 210 nm will reduce the selectivity of the detection, but analysis of an MSA blank (Figure 1) reveals that there are no interfering peaks at the retention times of *N,N*-dimethyl-*o*-toluidine and *N,N*-diethyl-*p*-toluidine. The spectral scans were added to a library used for confirming the presence of these amines in samples.




Figure 2. Separation of three mixed standard solutions of *N,N*-dimethyl-*o*-toluidine and *N,N*-diethyl-*p*-toluidine (1, 5, and 10 mg/L each) using the method in Figure 1.




Figure 3. Spectral scan of *N,N*-dimethyl-*o*-toluidine (5 mg/L).




Figure 4. Spectral scan of *N,N*-diethyl-*p*-toluidine (5 mg/L).

Table 3. Method Calibration Results as Reported by Chromeleon				
Peak Name	Coeff. Det. %	R-Square %	Offset	Slope
<i>N,N</i> -dimethyl- <i>o</i> -toluidine	99.9911	99.9910	-0.6125	8.9901
<i>N,N</i> -diethyl- <i>p</i> -toluidine	99.9880	99.9880	-1.3458	8.3587

Prior to sample analysis the system was calibrated using single injections of three concentrations of the two amines shown in Table 2 and Figure 2. Calibration (linear) results are shown in Table 3. Fourteen samples were analyzed, seven scrubbed ethylene gas samples and their seven corresponding blanks. *N,N*-diethyl-*p*-toluidine was not detected in any of the samples, but three samples had *N,N*-dimethyl-*o*-toluidine, all below the lowest concentration of the calibration standards. Figure 5 shows a chromatogram of one of the samples that had *N,N*-dimethyl-*o*-toluidine. Its presence was confirmed by a spectral match with the library.

This application update shows that IC columns designed for determining amines with suppressed conductivity and IPAD can also be used with MSA gradients and UV absorbance detection to determine amines not commonly determined by IC.




Figure 5. Chromatogram of the extract (scrubbed) of an ethylene gas sample using the method in Figure 1.

REFERENCES

1. Dionex Corporation. *Determination of Cations and Amines in Hydrogen Peroxide by Ion Chromatography Using a RFIC™ (Reagent-Free) System*. Application Update 155; LPN 1832. Sunnyvale, CA, 2006.
2. Dionex Corporation. *Determination of Biogenic Amines in Alcoholic Beverages by Ion Chromatography with Suppressed Conductivity and Integrated Pulsed Amperometric Detection*. Application Note 182; LPN 1888. Sunnyvale, CA, 2007.
3. Dionex Corporation. *Determination of Biogenic Amines in Fermented and Non-Fermented Foods Using Ion Chromatography with Suppressed Conductivity and Integrated Pulsed Amperometric Detection*. Application Note 183; LPN 1889. Sunnyvale, CA, 2007.

RFIC and Reagent-Free are trademarks and Chromeleon and IonPac are registered trademarks of Dionex Corporation.

Passion. Power. Productivity.



Dionex Corporation

1228 Titan Way
P.O. Box 3603
Sunnyvale, CA
94088-3603
(408) 737-0700

North America

U.S. (847) 295-7500
Canada (905) 844-9650

South America

Brazil (55) 11 3731 5140

Europe

Austria (43) 1 616 51 25 Benelux (31) 20 683 9768 (32) 3 353 4294
Denmark (45) 36 36 90 90 France (33) 1 39 30 01 10 Germany (49) 6126 991 0
Ireland (353) 1 644 0064 Italy (39) 02 51 62 1267 Switzerland (41) 62 205 9966
United Kingdom (44) 1276 691722

Asia Pacific

Australia (61) 2 9420 5233 China (852) 2428 3282 India (91) 22 2764 2735
Japan (81) 6 6885 1213 Korea (82) 2 2653 2580 Singapore (65) 6289 1190

www.dionex.com

LPN 1897 PDF 09/16
©2016 Dionex Corporation