

Efficient and Fast Separations of Inorganic Anions in Water Samples Using a 4 μm Particle Size Microbore Column with a High-Pressure Ion Chromatography System

Carl Fisher, Terri Christison, Barbara Shao, and Linda Lopez
Thermo Fisher Scientific, Sunnyvale, Ca, U.S.A.

Key Words

Drinking water, Wastewater, Municipalities, IonPac AS18-4 μm , HPIC, ICS-5000+

Goal

Demonstrate high resolution and short run times using a 4 μm resin anion-exchange separation column on a high-pressure ion chromatography system.

Introduction

In 1993, the U.S. EPA Method 300.0 (A) defined the use of the Thermo Scientific™ Dionex™ IonPac™ AS4A anion-exchange column to determine inorganic anions in environmental waters using a manually prepared carbonate/bicarbonate eluent and suppressed conductivity detection.¹ With the introduction of Reagent-Free™ ion chromatography (RFIC™) systems, hydroxide-based eluents were electrolytically generated inline on anion-exchange columns optimized for hydroxide chemistry. Reagent-free IC systems deliver reproducible and precise eluent concentrations that are produced inline by the addition of just water. These systems eliminate labor that would normally be spent on manual eluent preparation, removing a potential source of error and operator-to-operator variability.

The Dionex IonPac AS18-4 μm column is a newly developed high capacity column with AS18 chemistry using 4 μm resin particles.² The Dionex IonPac AS18 column was developed as an alternative to the AS4A column; the AS18 is designed for use with EPA Method 300.0 (A) using Electrolytically Generated hydroxide eluents. These smaller resin particles allow an optimal combination of fast speed and high resolution. For the determination of inorganic anions in water samples, a comparison of the new 4 μm versus the previous 7.5 μm resin Dionex IonPac AS18 column is performed.



With the introduction of high-pressure IC, the Thermo Scientific™ Dionex™ ICS-5000+ Reagent-Free HPIC™ system has been designed to operate at pressures up to 5000 psi using all column formats: standard (4 mm i.d.), microbore (2 mm i.d.), and capillary (0.4 mm i.d.). Higher operating pressures allow the use of faster flow rates, which result in shorter run times.

This document presents the advantages of using the small-particle size microbore (2 mm i.d.) Dionex IonPac AS18-4 μm column, in combination with the Dionex ICS-5000+ HPIC system to obtain fast and efficient separation of inorganic anions in municipal drinking and wastewater samples.

Equipment

Dionex ICS-5000+ HPIC system

- Thermo Scientific Dionex SP/DP Pump module
- Thermo Scientific Dionex EG Eluent Generator module with high-pressure degas module
- Thermo Scientific Dionex DC Detector/Chromatography module
- Thermo Scientific Dionex AS-AP Autosampler
- Thermo Scientific™ Dionex™ Chromeleon™ Chromatography Data system (CDS) software ver. 6.8 or 7.1

Reagents and Standards

18 MΩ-cm degassed deionized water

Fisher Scientific reagents, ACS Grade

Samples

Municipal drinking and wastewater samples collected from two cities in California.

Conditions

Columns:	Dionex IonPac AG18-4μm, 2 × 50 mm Dionex IonPac AS18-4μm, 2 × 150 mm
Eluent Source:	Thermo Scientific Dionex EGC 500 KOH Eluent Generator Cartridge Thermo Scientific Dionex CR-ATC 500 Continuously Regenerated Anion Trap Column
Eluent:	23 mM KOH
Flow Rate:	0.25, 0.30, 0.38 mL/min for flow rate experiments 0.38 mL/min for samples
Column Temp.:	30 °C
Inj. Volume:	5 μL
Detection:	Suppressed conductivity, Thermo Scientific™ Dionex™ ASRS™ 300 Self-Regenerating Suppressor, recycle; 15, 18, 22 mA
Background Conductance:	< 1.0 μS
Noise:	< 3 nS
System backpressure:	2500–3600 psi




Figure 1. Flow diagram for the high-pressure Dionex ICS-5000+ Reagent-Free HPIC system.

Table 1. Consumables Table.

Product name	Type, Capillary	Dionex Part Number
Dionex EGC 500 KOH*	Anion Eluent Generator cartridge	075778
Dionex CR-ATC 500*	Anion electrolytic trap column	075550
Dionex IonPac AG18-4μm	Anion guard column	076037
Dionex IonPac AS18-4μm	Anion separation column	076036
Dionex ASRS 300*	Anion suppressor	064555
Dionex HP Degas Module*	Degas module	075522
Dionex HP fittings (blue)	Bolts / Ferrules*	074449 / 074373
EG Degas HP cartridge	High pressure degas cartridge, up to 5000 psi	074459
Dionex AS-AP autosampler vials	Package of 100, polystyrene vials, caps, blue septa	074228

Standard and Sample Preparation

The calibration standard solutions were prepared by diluting individual standard stock solutions to obtain the final concentrations listed in Figures 2–4.

Sample solutions were diluted with 18 MΩ-cm resistivity deionized water and filtered with a syringe filter (0.45 μm) without any pretreatment prior to injection. The amount of sample dilution is indicated in Figure 4.

Instrument Setup and Installation

Tip: To achieve the best chromatography with high-pressure IC, it is important to use high pressure connectors and ferrules (see Table 1) for all connections prior to the suppressor. The high-pressure Dionex ICS-5000⁺ HPIC Reagent-Free system is designed to operate from 2000 to 5000 psi. To set up this application, plumb the consumables and modules of this system according to Figure 1.

Install and hydrate the Dionex EGC 500 cartridge, Dionex CR-TC 500 trap column, and 2 mm Dionex SRS 300 suppressor according to the product manual instructions. Complete the installation according to the flow diagram, using high-pressure connectors and ferrules (blue) for all of the fittings from the pump prior to the suppressor. Standard pressure fittings can be used for the suppressor and detector connections. Detailed instructions are described in Technical Note TN 129, the product manuals, and the instrument installation and operator's manuals.^{3–6}

Results and Discussion

Figure 2 shows a comparison of the smaller-particle-size (4 μm) column Dionex IonPac AS18-4μm, 2 × 150 mm with the larger-particle-size (7.5 μm) column Dionex IonPac AS18-Fast, 7.5 μm, 2 × 150 mm for the separation of nine anions in a standards mixture. Note that only half the amount of standard mixture was injected onto the 4 μm column. Good separation was obtained with both columns, although resolution and peak efficiency were enhanced with the smaller-particle-size column (4 μm). This better resolution can be exploited to achieve faster run times by using elevated flow rates without the loss of peak discrimination.

Figure 3 illustrates the analysis of nine inorganic anions using a Dionex IonPac AS18-4μm, 2 × 150 mm column with several flow rates. At a flow rate of 0.38 mL/min, all of the analytes were separated within 6 minutes with no loss of resolution compared to the slower runs. At this flow rate, a backpressure of 3600 psi is produced, which is well within the maximum pressure limit (5000 psi) of the Dionex ICS-5000⁺ HPIC system.

Column:	A: Dionex IonPac AS18-Fast, 2 × 150 mm B: Dionex IonPac AS18-4μm, 2 × 150 mm		
Eluent Source:	A: Dionex EGC III KOH B: Dionex EGC 500 KOH		
Eluent:	23 mM KOH		
Flow Rate:	0.25 mL/min		
Inj. Volume:	A: 5 μL; B: 2.5 μL		
Column Temp.:	30 °C		
Detection:	Suppressed conductivity, Dionex ASRS 300, 2 mm, recycle, 15 mA		

Peaks:	1. Fluoride	0.5 mg/L	5. Carbonate	20.0 mg/L
	2. Chlorite	5.0	6. Bromide	10.0
	3. Chloride	3.0	7. Sulfate	10.0
	4. Nitrite	5.0	8. Nitrate	10.0
			9. Chlorate	10.0




Figure 2. Comparison of the separation of nine inorganic anions using small-particle (4 μm) and large-particle (7.5 μm) analytical columns.

Column:	Dionex IonPac AG18-4μm, Dionex IonPac AS18-4μm, 2 × 150 mm		
Eluent Source:	Dionex EGC 500 KOH		
Eluent:	23 mM KOH		
Flow Rate:	A: 0.25, B: 0.30, C: 0.38 mL/min		
Inj. Volume:	5 μL		
Column Temp.:	30 °C		
Detection:	Suppressed conductivity, Dionex ASRS 300, 2 mm, recycle		
	A: 15, B: 18, C: 22 mA		
Back Pressure:	A: 2500; B: 3000; C: 3600 psi		

Peaks:	1. Fluoride	0.5 mg/L	6. Bromide	10.0 mg/L
	2. Chlorite	1.0	7. Sulfate	10.0
	3. Chloride	3.0	8. Nitrate	20.0
	4. Nitrite	5.0	9. Chlorate	10.0
	5. Carbonate	--		




Figure 3. Comparison of the separation of inorganic anions using different flow rates.

Figure 4 shows the determination of inorganic anions in municipal drinking and waste water samples using a Dionex IonPac AS18-4 μ m, 2 $\times$ 150 mm column. The five anions that are commonly found in municipal water samples were resolved within 5 min, demonstrating the speed with which such runs can be performed using this column.




Figure 4. Determination of inorganic anions in municipal drinking and waste water samples.

Conclusion

This technical note demonstrates how the Dionex IonPac AS18-4 μ m anion-exchange column can be used in combination with the Dionex ICS-5000+ HPIC system to obtain efficient separations with high resolution. It was demonstrated that all five anions present in samples eluted within 6 minutes. The efficiency of the Dionex IonPac AS18-4 μ m analytical column permits the use of higher flow rates, reduces run times and, therefore, increases throughput for determination of inorganic anions in environmental water samples.

Additional information on inorganic anion determination in environmental and drinking water samples using standard bore and microbore IC are reviewed in application note AN 154 and application brief AB 132.⁷⁻⁸

References

1. U.S. EPA Method 300.0, rev 2.1, Determination of Inorganic Anions in Water by Ion Chromatography, U.S. Environmental protection Agency, Cincinnati, OH, 1993.
2. Thermo Scientific Dionex IonPac AS18-4 μ m Hydroxide-Selective Anion-Exchange Capillary Column, <http://www.dionex.com/en-us/products/columns/ic-rfic/hydroxide-selective-packed/ionpac-as18-4um/lp-111535.html>
3. Thermo Fisher Scientific. Dionex Technical Note 129, Configuring High-Pressure IC Systems for Analytical Flow Rates, Dionex Doc No. TN70317, Sunnyvale, CA, 2012.
4. Thermo Fisher Scientific. Dionex Product Manual for Anion Self-Regenerating Suppressor 300 and Cation Self-Regenerating Suppressor 300. Dionex Doc No. 031956, Sunnyvale, CA, 2009.
5. Thermo Fisher Scientific. ICS-5000+ Installation instructions. Dionex Doc No. 065447, Sunnyvale, CA, 2011.
6. Thermo Fisher Scientific. Dionex AS-AP Operator's Manual. Dionex Doc No. 065361, Sunnyvale, CA, 2012.
7. Thermo Fisher Scientific. Application Note 154, *Determination of Inorganic Anions in Environmental Water Using a Hydroxide-Selective Column*. Dionex LPN 1539, Sunnyvale, CA 2003.
8. Thermo Fisher Scientific. Application Brief 132, *Determination of Anions in Municipal Drinking Water by Fast IC Using a Hydroxide Eluent*. Dionex LPN 2874, Sunnyvale, CA 2011.

www.thermofisher.com/dionex

©2016 Thermo Fisher Scientific Inc. All rights reserved. PEEK is a trademark of Victrex PLC. All other trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries. This information is presented as an example of the capabilities of Thermo Fisher Scientific Inc. products. It is not intended to encourage use of these products in any manners that might infringe the intellectual property rights of others. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

Australia +61 3 9757 4486
Austria +43 1 333 50 34 0
Belgium +32 53 73 42 41
Brazil +55 11 3731 5140
China +852 2428 3282

Denmark +45 70 23 62 60
France +33 1 60 92 48 00
Germany +49 6126 991 0
India +91 22 2764 2735
Italy +39 02 51 62 1267

Japan +81 6 6885 1213
Korea +82 2 3420 8600
Netherlands +31 76 579 55 55
Singapore +65 6289 1190
Sweden +46 8 473 3380

Switzerland +41 62 205 9966
Taiwan +886 2 8751 6655
UK/Ireland +44 1442 233555
USA and Canada +847 295 7500

Thermo
 SCIENTIFIC
 Part of Thermo Fisher Scientific