

# Characterization of Unsaturated Fatty Acids in Negative OAD-MS/MS using LCMS-9050

Hidenori Takahashi, Mami Okamoto, Yuta Miyazaki, Yohei Arao, Natsuyo Asano  
Shimadzu Corporation, Kyoto, Japan

## 1. Introduction to OAD-MS/MS

- While low-energy CID-MS/MS is one of the most effective fragmentation techniques for structural analysis, it may not be ideally suited for the analysis of certain isomers.
- Several novel fragmentation techniques have been proposed to complement low-energy CID-MS/MS.

**Table 1. Example of proposed novel fragmentation techniques**

### Electron-based fragmentation

EIEIO, ECD(Electron Capture Dissociation) by Zubarev et al. (1996)

### Anion-based fragmentation

ETD(Electron Transfer Dissociation) by Syka et al. (2004)

### Photon-based fragmentation

IRMPD (Infrared), UVPD (Ultraviolet), BRID (Blank body infrared)


- We have introduced **neutral radical-based** fragmentation techniques to structural analysis of biomolecules, peptides and lipids, since 2016.

**Neutral radical-based** \*Takahashi et al., *Anal. Chem.* 2018, 90, 12, 7230.

Charge-neutral radical-induced dissociation is available in both positive and negative ion modes!

**O•** OAD (Oxygen Attachment Dissociation)-MS/MS  
 $[M+H]^+ + O^{\bullet} \rightarrow [M+H+O]^+ \rightarrow \text{fragments}$


**H•** HAD (Hydrogen Abstraction Dissociation)-MS/MS  
 $[M+H]^+ + H^{\bullet} \rightarrow M^{\bullet} + H_2 \rightarrow \text{fragments}$



**Fig. 1. Shimadzu LCMS-9050 (Q-TOF) with OAD unit.**


## 2. Lipid MS/MS: CID vs. OAD vs. HAD Comparison

- CID-MS/MS** does not provide detailed structural information within carbon chains. Instead, CID selectively cleaves labile polar head groups ( $m/z$  184).




**Fig. 2-1. Typical CID spectrum of a model lipid of PC (18:1).**

- OAD-MS/MS** clearly provides C=C positional information. Atomic oxygen selectively oxidizes and cleaves at C=C.



**Fig. 2-2. Typical OAD spectrum of a model lipid of PC (18:1).**

- HAD-MS/MS** provides sequential structural information within carbon chains. The sensitivity of HAD is lower than OAD for C=C position assignment, since fragment ions become too complicated in a complex sample mixture.



**Fig. 2-3. Typical HAD spectrum of a model lipid of PC (18:1).**

## 3. Negative OAD-MS/MS in lipid analysis

**Table 2. Analytical Condition**

### HPLC condition

**System:** Shimadzu Nexera X3

**Flow rate:** 0.3 mL/min

**Mobile phase A:** 1:1:3 (v/v/v) ACN:MeOH: water with ammonium acetate (5mM) and 10nM EDTA

**Mobile phase B:** 100% IPA with ammonium acetate (5mM) and 10nM EDTA.

**Column:** Acquity UPLC Peptide BEH C18 (50 × 2.1mm; 1.7μm; Waters, USA)

**Gradient:** 0min 0% (B); 1min 0% (B); 5min 40% (B); 7.5min 64% (B); 12min 64% (B); 12.5min 82.5% (B); 19 min 85% (B); 20min 95% (B); 20.1min 0% (B); and 25min 0% (B).

### MS condition

System: Shimadzu LCMS-9050



Mode: OAD

Polarity: Negative ESI

CE: +10 V


MS/MS: Auto-MS/MS Top 10

**Sample:** Lipid extract from a mouse's brain with the addition of several commercially available fatty acid standards. The method for preparing the lipid extract adheres to the procedure described in Uchino.H et al. *Commun Chem.* 5, 162 (2022).




**Fig. 3. Example of negative-OAD spectrum at several retention time.**

### FA 18:2(n-6, 9) Linoleic acid



### PE 18:0\_20:4(n-6, 9, 12, 15)



### Co-eluted isomers


#### FA 18:1(n-9) Oleic acid



#### FA 18:1(n-7) Vaccenic acid



### PI 18:0\_22:6(n-3, 6, 9, 12, 15, 18)



## 4. Conclusions

- OAD was successfully integrated into the QTOF of LCMS-9050.
- OAD has demonstrated its effectiveness as a powerful tool for negative MS/MS fragmentation.
- Without any need for derivatization, OAD can accurately assign the C=C bond positions in lipids and fatty acids.

