

## Application News

# No.**C100**

Liquid Chromatography Mass Spectrometry

### Quantitative Analysis of Pyrethroids in Soil and Sediment Using the Shimadzu LCMS-8050 Triple Quadrupole Mass Spectrometer

Pyrethroid pesticides are used widely around the world as agricultural and household insecticides. Synthetic pyrethroids are slightly soluble in water and easily adsorbed in soil. In recent years, pyrethroid residues have been confirmed in soil and sediment in both agricultural and urban areas. Pyrethroids, while posing little danger to humans, exhibit a high toxicity to aquatic organisms and insects, making their impact on the ecosystem a matter of concern. Therefore, there is

#### a need for a sensitive technique which can rapidly measure pyrethroid pesticides in soil and sediment. Due to their low polarity, pyrethroid pesticides are typically measured by GC and GC-MS, however this Application News demonstrates simultaneous positiveand negative-ion mode analysis of 14 pyrethroid pesticides using LC-MS/MS with electrospray ionization (ESI).

#### MRM Analysis of Standards and Generation of Calibration Curves

1500000 1400000 1300000-1200000 1100000pyrethrin-II 1000000pyrethrin-I 900000月 fentropathrin 800000cycloprothrin 700000 deltamethrin 600000esfenvalerate cypermethrin 500000cyfluthrin 400000ethofenprox 300000 F permethrin 200000cyhalothrin bifenthrin 100000 acrinathrin 0 silafluofen 7.0 8.0 9.0 10.0 min

Fig. 1 MRM Chromatograms of Pyrethroids

Table 1 MRM Transitions and Calibration Curves of Pyrethroids

Compound Name	Polarity	Precursor ( <i>m/z</i> )	Product ( <i>m/z</i> )	Calibration Curve Range (µg/L)	r <sup>2</sup>
Pyrethrin- I	+	329.20	161.10	0.5 - 500	0.9996
Pyrethrin- II	+	373.20	161.20	0.5 - 500	0.9997
Fenpropathrin	+	367.20	125.20	0.02 - 100	0.9993
Cycloprothrin	+	498.90	181.10	0.5 - 100	0.9991
Deltamethrin	+	522.80	280.90	0.05 - 100	0.9992
Esfenvalerate	+	437.10	167.30	0.5 - 100	0.9990
Cypermethrin	+	433.10	191.10	0.05 - 100	0.9986
Cyfluthrin	+	450.90	191.00	0.5 - 100	0.9976
Ethofenprox	+	394.20	177.30	0.01 - 100	0.9993
trans-Permethri	in +	408.10	183.30	0.02 - 100	0.9996
cis-Permethrin	+	408.10	183.30	0.02 - 100	0.9994
Cyhalothrin	+	467.10	225.10	0.1 - 100	0.9993
Bifenthrin	+	440.00	181.20	0.02 - 100	0.9995
Acrinathrin	-	540.10	372.20	0.1 - 100	0.9994
Silafluofen	+	426.20	287.10	0.01 - 100	0.9999




Fig. 2 MRM Chromatograms at Pyrethroid LOQs

#### Sample Preparation for Soil and Sediment Using QuEChERS Method

Soil samples are generally prepared using solid-phase extraction, however, this process can be both timeconsuming and labor-intensive. In this application, the easy pretreatment method referred to as QuEChERS (Quick, Easy, Cheap, Effective, Rugged, Safe), typically used for analysis of residual pesticides in food, was used for the pretreatment of soil and sediment samples. Fig. 3 shows the protocol employed to pretreat these soil and sediment samples. The combined acetonitrile extraction and cleanup process requires only 15 minutes per sample to complete.

#### Recoveries Using Actual Samples

A mixed pesticide standard solution was added to soil and sediment solutions respectively to obtain a concentration of 10 ppb before or after pretreatment was conducted by the QuEChERS method, and recovery tests were then conducted. Good recoveries of 70 to 120 % were obtained for both soil and sediment samples, as shown in Fig. 4.




Fig. 3 Sample Preparation Using QuEChERS Method




Fig. 4 Recoveries from Soil and Sediment

#### Table 2 Analytical Conditions

Column	: Phenomenex Kinetex 2.6 µm PFP 100Å (100 mm × 2.1 mm I.D.)			
Mobile Phase A	: 5 mM Ammonium Acetate - Water			
Mobile Phase B	: Methanol			
Time Program	: 40 %B (0 min) $\rightarrow$ 100 %B (10 - 12 min) $\rightarrow$ 40 %B (12.01 - 15 min)			
Flowrate	: 0.2 mL/min.			
Injection Volume	: 1 µL			
Oven Temperature	: 40 °C			
Ionization Mode	: ESI (Positive / Negative)			
Probe Voltage	: +4.0 kV / -3.0 kV			
Neburizing Gas Flow	: 3.0 L/min.			
Drying Gas Flow	: 15.0 L/min.			
Heating Gas Flow	: 15.0 L/min.			
Interface Temperature	: 100 °C			
DL Temperature	: 100 °C			
Block Heater Temperature : 400 °C				

🕀 SHIMADZU

Shimadzu Corporation www.shimadzu.com/an/

For Research Use Only. Not for use in diagnostic procedures.

The content of this publication shall not be reproduced, altered or sold for any commercial purpose without the written approval of Shimadzu. The information contained herein is provided to you "as is" without warranty of any kind including without limitation warranties as to its accuracy or completeness. Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication. This publication is based upon the information available to Shimadzu on or before the date of publication, and subject to change without notice.

First Edition: Jan. 2015