

NOTICE: Varian, Inc. was acquired by Agilent Technologies in May 2010. This document is provided as a courtesy but is no longer kept current and thus will contain historical references to Varian. For more information, go to [www.agilent.com/chem](http://www.agilent.com/chem).



## Application Note SI-01984

# High Temperature SimDist Analysis According to IP507/07

Coen Duvekot & Klaas Oostdijk  
Varian, Inc.

### Introduction

The IP507/07 standard specifies a method for the determination of the boiling range distribution of petroleum products by capillary gas chromatography using flame ionization detection. The standard is applicable to materials having a vapor pressure low enough to permit sampling at ambient temperature and a boiling range of at least 100 °C. The standard is applicable to materials with initial boiling points (IBP) above 100 °C and final boiling points (FBP) above 750 °C, for example, heavy distillate fuels and residuals. The method is not applicable to bituminous samples.

### Instrumentation

GC: Varian 450-GC/Simulated Distillation Analyzer  
Injector: Temperature controlled with Varian SPI injector for capillary column with full EFC control  
Column oven: With cryogenic (CO<sub>2</sub>) cooling  
Detector: FID with EFC control  
Autosampler: CP-8400 or CP-8410

### Software

GC control and data handling: Galaxie™ GC Workstation  
SimDist calculations: SimDist software plug-in fully integrated into Galaxie

### Materials

Column: Varian CP-SimDist column, 5 m x 0.53 mm x 0.1 μm  
Ultimetel (pn: CP7569)  
Calibration mix for determination of TBP range  
Motor oil for 100% recovery determination

### Sample Preparation

All samples and calibration mixtures are dissolved in CS<sub>2</sub> at 2%.

### Conditions

Oven temp/program: 35 °C/10 °C per min to 430 °C  
Injector temp/program: 100 °C/15 °C per min to 430 °C  
Detector temp: 450 °C  
Carrier gas/flow rate: Helium, 19 mL/min  
Sample size: 1 μL

### Results




Figure 1. Typical chromatogram of a calibration mix.




Figure 2. Typical calibration curve.




Figure 3. Chromatogram of a motor oil sample. Recovery is 100%.




Figure 4. Result plot of the motor oil sample.




Figure 5. Chromatogram of a residue sample. Recovery is 88.7%.

Table 1. Results from the residue sample.

% off	°C	D86
IBP	476.7	504.8
1	487.8	
5	520.6	528.8
10	539.3	548.9
15	553.9	
20	566.9	572.6
25	578.1	
30	587.5	587.5
35	596.0	
40	605.8	
45	615.4	
50	626.1	613.3
55	637.5	
60	649.7	
65	662.4	
70	676.3	652.9
75	692.0	
80	709.1	537.6
85	728.8	
88	745.0	

### Conclusion

The Varian SimDist Analyzer and its Galaxie™ based software provide the solution for high temperature simulated distillation applications as specified in SimDist method IP507/07.

### Reference

IP 507: Determination of boiling range distribution by gas chromatography method - Part 2: Heavy distillates and residual fuels. Energy Institute, London.

These data represent typical results.

For further information, contact your local Varian Sales Office.

Varian and the Varian Logo are trademarks or registered trademarks of Varian, Inc. in the U.S. and other countries.

© 2009 Varian, Inc.

Application Note SI-01984

Varian, Inc.  
[www.varianinc.com](http://www.varianinc.com)  
 North America: 800.926.3000 – 925.939.2400  
 Europe: *The Netherlands*: 31.118.67.1000  
 Asia Pacific: *Australia*: 613.9560.7133  
 Latin America: *Brazil*: 55.11.3238.0400



**VARIAN**